#### §868.5270

(b) Classification. Class II (performance standards).

# §868.5270 Breathing system heater.

- (a) *Identification*. A breathing system heater is a device that is intended to warm breathing gases before they enter a patient's airway. The device may include a temperature controller.
- (b) Classification. Class II (performance standards).

## §868.5280 Breathing tube support.

- (a) *Identification*. A breathing tube support is a device that is intended to support and anchor a patient's breathing tube(s).
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §868.9.
- [47 FR 31142, July 16, 1982, as amended at 54 FR 25048, June 12, 1989; 66 FR 38794, July 25, 2001]

#### §868.5300 Carbon dioxide absorbent.

- (a) *Identification*. A carbon dioxide absorbent is a device intended for medical purposes that consists of an absorbent material (e.g., soda lime) that is intended to remove carbon dioxide from the gases in the breathing circuit.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §868.9.
- [47 FR 31142, July 16, 1982, as amended at 61 FR 1120, Jan. 16, 1996; 66 FR 38794, July 25, 2001]

#### §868.5310 Carbon dioxide absorber.

- (a) *Identification*. A carbon dioxide absorber is a device that is intended for medical purposes and that is used in a breathing circuit as a container for carbon dioxide absorbent. It may include a canister and water drain.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §868.9.
- $[47\ FR\ 31142,\ July\ 16,\ 1982,\ as\ amended\ at\ 61\ FR\ 1120,\ Jan.\ 16,\ 1996;\ 66\ FR\ 38794,\ July\ 25,\ 2001]$

#### §868.5320 Reservoir bag.

- (a) *Identification*. A reservoir bag is a device, usually made of conductive rubber, intended for use in a breathing circuit as a reservoir for breathing gas and to assist, control, or monitor a patient's ventilation.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §868.9.
- [47 FR 31142, July 16, 1982, as amended at 61 FR 1120, Jan. 16, 1996; 66 FR 38794, July 25, 2001]

#### §868.5330 Breathing gas mixer.

- (a) *Identification*. A breathing gas mixer is a device intended for use in conjunction with a respiratory support apparatus to control the mixing of gases that are to be breathed by a patient.
- (b) Classification. Class II (performance standards).

### §868.5340 Nasal oxygen cannula.

- (a) *Identification*. A nasal oxygen cannula is a two-pronged device used to administer oxygen to a patient through both nostrils.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §868.9.
- [47 FR 31142, July 16, 1982, as amended at 59 FR 63007, Dec. 7, 1994; 66 FR 38794, July 25, 2001]

## §868.5350 Nasal oxygen catheter.

- (a) *Identification*. A nasal oxygen catheter is a device intended to be inserted through a patient's nostril to administer oxygen.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §868.9.
- [47 FR 31142, July 16, 1982, as amended at 59 FR 63007, Dec. 7, 1994; 66 FR 38794, July 25, 2001]

# §868.5365 Posture chair for cardiac or pulmonary treatment.

(a) Identification. A posture chair for cardiac or pulmonary treatment is a